

*Official copies of these procedures are maintained at this website.
Before using a printed copy, verify that it is the most current
version by checking the document issue date on this website. Signed
copies of these official procedures are maintained at the Training Office*

C-A OPERATIONS PROCEDURES MANUAL

ATTACHMENT

4.61.a NMC Recheck Sheet

Text page 2

C-A OPM Procedures in which this Attachment is used.
--

4.61		

Hand Processed Changes

<u>HPC No.</u>	<u>Date</u>	<u>Page Nos.</u>	<u>Initials</u>
_____	_____	_____	_____
_____	_____	_____	_____
_____	_____	_____	_____

Approved: _____ **Signature On File** _____
Collider-Accelerator Department Chairman Date

D. Gassner

C-A-OPM ATT 4.61.a

1

Revision 00
August 16, 2000

NMC RECHECK DATA SHEET

Upon commissioning and at monthly intervals during operations of the applicable beam line, NMC monitors will have the following data recorded in a special log:

1. Date _____
2. By _____
3. S/N of NMC unit _____
4. S/N of paddle _____
5. Upscale Check voltage at input to integrator card _____
6. Voltage measured using a Fluke model 23 DVM (from TB1-5)

LAMP VOLTS TB1-4	METER READING	CONTROL VOLTS card pin M	HIGH VOLTS TB1-3
0V			
2V			
4V			
6V			

FINAL SETUP
